

**GOVERNMENT OF RAJASTHAN
DIRECTORATE MEDICAL & HEALTH SERVICES RAJASTHAN**

No. SPO/E&I/App. List/2008-10/09/ 1038

Dated: 20-3-09

All Administrative Officers

Sub: - Purchase of Store & Approval of Items (Equipments & Instruments) Part-B, Section-C (Anesthetic and Operation Theatre Equipment & Instrument,) Cat No. C/1 (Adult laryngoscope) C/2 (Paed. laryngoscope) C/6- (Bulk conversion) C/16 (Child silicon Resuscitator) C/17- (Infant silicon Resuscitator), C/18- (Adult Bain Circuit Complete) H/2- (Water Bath Serological) H/3- (Hot Air Oven) & H/14- (Incubator Size 400 x 400 x 400) for the Rate contract period 2008-10 ending on 30-09-2010.

In continuation to this Directorate letter no. SPO/E&I/App. List/2008-2010/08/2779 dated 22.10.08, No. 2799 dated: 27-10-08, 3004 dated 25-11-08, 3209 dated 31.12.08, 382 dated 28-01-09, 478 dated 04-02-09, 558 dated 09-02-09, 567 dated 10-02-09, 623 Dated 13-02-09. 720 Dated 20-02-09 & 987 dated 16.03.09. Please find enclosed herewith the approved list showing the details such of the firm and rates etc. approved in respect of the firm and rates approved in respect of item Cat No C/1 (Adult laryngoscope) C/2 (Paed. laryngoscope) C/6- (Bulk conversion) C/16 (Child silicon Resuscitator) C/17- (Infant silicon Resuscitator), C/18- (Adult Bain Circuit Complete) H/2- (Water Bath(Serological)) H/3- (Hot Air Oven) & H/14- (Incubator 400x400x400) for the Rate contract The terms & conditions will remain same as already communicate to you earlier vide letter No. SPO/E&I/App. List/2008-2010/08/2779 Dated: 22-10-08.

The supplier has executed the agreement bond. Therefore the payment of the supplier should be released to them, provided the goods are according to approved specification and rates, to your entire satisfaction as per rules.

It may please be ensured, if there is any recovery against the approved suppliers pertaining to the previous/ current rate contract on account of Liquidated Damages/ Risk Purchases/ Breakage/ Shortage/ Sub Standard etc. the same may please be recovered at your end before making payment.

Receipt of this communication may please be acknowledged.

Encl: - As above

Director (PH)
Medical & Health Services
Rajasthan, Jaipur

Dated: 20-3-09

No. SPO/E&I/App. List/2008-10/09/ 1038

Copy forwarded to the following for information & N/A:

1. PA To DM&HS (PH)/ (Aids)/(FW).
2. PA To Project Director, RHSDP/ NRHM.
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota & Udaipur.
4. Superintendent, SMS Hospital, Zanana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, TB & Chest Hospital, Jaipur.
5. Superintendent Associated Group Of Hospital-Ajmer, M.G. Hospital- Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital- Kota, & A.G. Hospital-Udaipur.
6. Director, Medical & Health Services (EST), Mobile Surgical Unit, Raj., Jaipur.
7. Addl. Director, Medical & Health Services (FW)/ (RH)/ (Hospital Admn.).
8. PA to FA of This Directorate.
9. Joint Director, Medical & Health Services, (Zones) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
10. All In-charge T.B Clinic's
11. All Dy. CM&HO's
12. In-charge Server Room, Head Quarter
13. Concerned Firms.
14. Office Order File.

Chief Accounts Officer & Secretary,
Store Purchase Organisation,
Medical & Health Directorate,
Rajasthan Jaipur

**GOVERNMENT OF RAJASTHAN
DIRECTORATE MEDICAL, HEALTH F.W. SERVICES RAJASTHAN, JAIPUR**

Item and rate approved for the rate contract period 2008-10 ending 30.9.2010 for Part-B (Equipments & Instruments) Section-C (Anesthetic and Operation Theatre Equipment & Instrument)

R.C. No. XV, (Page No.3)

R.C. No: DMHS/SPO/E&I/RC-I/2008-10/09/ 1032

Dated: 20-3-09

S. No.	Cat. No.	Name of the Item with Specifications	Name of Firm with Address	Packing Unit	Brand	Approved Rate
1	2	3	4	5	6	7
1	C/1	Adult laryngoscope with handle and blade (Mackintosh type) S.S. Size : 1,2,3 & 4	M/s G.S. Healthcare Products Pvt. Ltd., 236/48, Biren Roy Road (West), Shibrampur (Sarada Pard) Kolkata. Tel. No. 033-24882103 Fax No. 033-24882693	Each	HCP Own Make	Rs. 900.00 (Rs. Nine hundred only) + 4% Tax extra or as applicable
2.	C/2	Paed. laryngoscope with handle and blade (Miller type) S.S. Size: 0 & 1	M/s G.S. Healthcare Products Pvt. Ltd., 236/48, Biren Roy Road (West), Shibrampur (Sarada Pard) Kolkata. Tel. No. 033-24882103 Fax No. 033-24882693	Each	HCP Own Make	Rs. 495.00 (Rs. Four hundred ninety five only) + 4% Tax extra or as applicable
3.	C/6	Bulk conversion unit for nitrous oxide : Pre set regulator with pressure gauge 0-250 kgf/cm2 with self sealing valve Nitrous oxide & 5 meter antistatic high pressure blue color tubing fitted with one side matching probe Nitrous oxide and one side pin index adopter N2O. Tube clamping with ferule.	M/s G.S. Healthcare Products Pvt. Ltd., 236/48, Biren Roy Road (West), Shibrampur (Sarada Pard) Kolkata. Tel. No. 033-24882103 Fax No. 033-24882693	Each	HCP Own Make	Rs. 3500.00 (Rs. Three thousand five hundred only) + 4% Tax extra or as applicable



NEXT Page

4.	C/16	Child silicon Resuscitator : 500ml Silicon bag Non Rebreathing valve with 40cm of H2O pressure release, Mask Size no. 01 3600 swiveling patient connector PVC Oxygen Reservoir, 1.5mtr. PVC Oxygen enrichment Tubing.	M/s Bharat Surgical Co., Main Road, Madan Puri, Gurgaon Tel. No. 0124- 2320033 Fax No./ 0124 2118733.	Each	Duke Own Make	Rs. 670.00 (Rs. Six hundred seventy only) + 4% Tax extra or as applicable
5.	C/17	Infant silicon Resuscitator : 250ml Silicon bag Non Rebreathing valve with 40cm of H2O pressure release Mask Size no. 00, 3600 swiveling patient connector PVC Oxygen Reservoir, 1.5mtr. PVC Oxygen Tubing.	M/s Bharat Surgical Co., Main Road, Madan Puri, Gurgaon Tel. No. 0124- 2320033 Fax No./ 0124 2118733.	Each	Duke Own Make	Rs. 652.00 (Rs. Six hundred fifty two only) + 4% Tax extra or as applicable
6.	C/18	Adult Bain Circuit Complete	M/s Bharat Surgical Co., Main Road, Madan Puri, Gurgaon Tel. No. 0124- 2320033 Fax No./ 0124 2118733.	Each	Duke Own Make	Rs.571.00 (Rs. Five hundred seventy one only) + 4% Tax extra or as applicable

Director (PH)
Medical & Health Services
Rajasthan, Jaipur

15/3

**GOVERNMENT OF RAJASTHAN
DIRECTORATE MEDICAL, HEALTH F.W. SERVICES RAJASTHAN, JAIPUR**

Item and rate approved for the rate contract period 2008-10 ending 30.9.2010 for Part-B (Equipments & Instruments) Section-H (Laboratory Equipments & Appliances)

R.C. No. XV, (Page No.3)

R.C. No: DMHS/SPO/E&I/RC-XV/2008-10/09/ 1038

Dated: 20-3-09

S.No.	Cat. No.	Name of the Item with Specifications	Name of Firm with Address	Packing Unit	Brand	Approved Rate	Remarks
1	2	3	4	5	6	7	8
7.	H/2	Water Bath(Serological): - Stainless steel Chamber Adjustable Temperature (range 10-70°c) Slot for thermometer, Visual display of Temperature, Stainless steel top cover. Size of inner chamber 330x300x150 mm. for 4 racks, load 1.5 KW. The Equipment should be provided with a panel having a thermostat control Knob with essential accessories.	M/s Linco Scientific Instruments Chemicals (P) Ltd., 62, Industrial Estate, Ambala Cantt.-133006 Phone No. 0171-2699412, 2642216 Fax No. 0171-2643506	Each	Microsil ®, Model- SWB-2	Rs. 3,990.00 (Rupees Three thousand nine hundred ninety only)+4% Tax or as applicable	20% of the institutional actual requirement shall be purchased From M/s Linco Scientific Instruments Chemicals (P) Ltd. Ambala Cantt
8.	H/3	Hot Air Oven: - Internal size 45x45x45 cms, single door with 2 selves, door fitted on heavy brass cast and Chrome Plated hinges, Insulation minimum thickness 2" Glass Wool, Inner Chamber of anodized Stainless steel elements on three sides Temperature is Controlled by thermostat from 50-250°c +_1°c with essential accessories.	M/s Linco Scientific Instruments Chemicals (P) Ltd., 62, Industrial Estate, Ambala Cantt.-133006 Phone No. 0171-2699412, 2642216 Fax No. 0171-2643506	Each	Microsil ®, Model- OVS-3	Rs. 8,800.00 (Rupees Eight thousand eight hundred only) + 4% Tax (against certificate that machine is to be used in Hospital as Medical device.) or as applicable	20% of the institutional actual requirement shall be purchased From M/s Linco Scientific Instruments Chemicals (P) Ltd. Ambala Cantt
9.	H/14	Incubator :- Inner Chamber of anodized stainless steel sheet size 400x400x400. Temperature to be maintained at 37°c +_1, digital Temperature indicator, uniform Temperature on all shelves, interior chamber insulated door with inner glass.	M/s Linco Scientific Instruments Chemicals (P) Ltd., 62, Industrial Estate, Ambala Cantt.-133006 Phone No. 0171-2699412, 2642216 Fax No. 0171-2643506	Each	Microsil ®, Model- IBS-2	Rs. 11,400.00 (Rupees Eleven thousand four hundred only) + 4% Tax or as applicable	—

Director (PH)
Medical & Health Services
Rajasthan, Jaipur

R.C. No. XV
10/38

20/3/09