

GOVERNMENT OF RAJASTHAN
DIRECTORATE MEDICAL & HEALTH SERVICES RAJASTHAN

No. SPO/E&I/App. List/2008-2010/1801

Date: 4/6/09

All Administrative Officers

Sub: - Purchase of Store & Approval of Items Part-B (Equipments & Instruments)
Section-M (FBNC & NBC) Item- Bilirubinometers & Multi-Parameter Monitor
for the Rate Contract & Supply period 2008-10 ending on 30-09-2010.

In continuation to this Directorate letter no. SPO/E&I/App. List/2008-2010/08/2779 dated 22.10.08, No. 2799 dated: 27-10-08, 3004 dated 25-11-08, 3209 dated 31.12.08, 382 dated 28-01-09, 478 dated 04-02-09, 558 dated 09-02-09, 567 dated 10-02-09, 623 Dated 13-02-09, 720 Dated 20-02-09, 987 dated 16.03.09, 1037 dated 18.3.09, 1038 dated 20.3.09, 1112 Dated 25-03-09, 1157 dated 31-03-09, 1353 dated 16-04-09, 1380 dated 20.4.09, 1381 dated 20.4.09, 1461 dated 27-04-09 & 1571 dated 08-05-09. Please find enclosed herewith the approved list showing the details such of the firm and rates etc. approved in respect of the firm and rates approved in respect of item Section-M (FBNC & NBC) Item- Bilirubinometers & Multi-Parameter Monitor for the Rate contract the terms & conditions will remain same as already communicate to you earlier vide letter No. SPO/ E&I/App. List/ 2008-2010/ 08/ 2779 Dated 22-10-08 & Special Terms & Condition enclose herewith.

The supplier has executed the agreement bond. Therefore the payment of the supplier should be released to them, provided the goods are according to approved specification and rates, to your entire satisfaction as per rules.

It may please be ensured, if there is any recovery against the approved suppliers pertaining to the previous/ current rate contract on account of Liquidated Damages/ Risk Purchases/ Breakage/ Shortage/ Sub Standard etc. the same may please be recovered at your end before making payment.

Receipt of this communication may please be acknowledged.

Encl: - As above

Director (PH)
Medical & Health Services
Rajasthan, Jaipur

Date: 4/6/09

No. SPO/E&I/App. List/2008-10/09/1801

Copy forwarded to the following for information & N/A:

1. PA To DM&HS (PH)/ (Aids)/(FW).
2. PA To Project Director, RHSDP/ NRHM.
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota & Udaipur.
4. Superintendent, SMS Hospital, Zanana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, TB & Chest Hospital, Jaipur.
5. Superintendent Associated Group Of Hospital-Ajmer, M.G. Hospital- Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital- Kota, & A.G. Hospital-Udaipur.
6. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
7. Addl. Director, Medical & Health Services (FW)/(RH)/ (Hospital Admn.)
8. PA to FA of This Directorate along with 5 copies.
9. Joint Director, Medical & Health Services, (Zones) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
10. All In-charge TB Clinic's
11. All Dy. GM&HO's
12. In-charge Server Room, Head Quarter
13. Concerned Firms.
14. Office Order File.

Chief Accounts Officer & Secretary,
Store Purchase Organisation,
Medical & Health Directorate,
Rajasthan Jaipur

3/6 4/6/09

GOVERNMENT OF RAJASTHAN
DIRECTORATE MEDICAL, HEALTH & W. SERVICES RAJASTHAN, JAIPUR.

Item and rate approved for the rate contract period 2008-10 ending 30.9.2010 for Part-B (Equipments & Instruments) Section -M (FBNC & NBC) Item- Multi-Parameter Monitor

R.C. No. XXIV (Page No. 3)

R.C. No: DMHS/SPO/E&I/RC-XXIV/2008-10/09/ 180)

Dated: 4/6/09

S. No.	Name of the Item with Specifications	Name of Firm with Address	Packing Unit	Brand	Approved Rate Per Packing Unit	Remarks
1	2	3	4	5	6	7
1	<p>Multi-Parameter Monitor</p> <ul style="list-style-type: none"> Should have the facility of monitoring ECG, RR, SpO2, NIBP & Temp for Neonatal & Pediatric applications. Should have integrated colour TFT display of size 10" or more Should have facility of viewing at least 4 waveforms simultaneously Should have arrhythmia detection with alarm facility Must use branded Nellcor or Masimo pulse oximetry module with facility for display of Plethysmograph, Pulse strength & SpO2 values, H.R. 40 to 250 bpm Should have non - volatile Graphical & Tabular trend facility for at least 24 hrs Should operate independently on both mains and battery Inbuilt rechargeable battery facility should be more than 4 hrs Networking facility with central nursing station should be possible. Should have alarm limits with alarm levels and alarm indication (visual - colour coded as well as audio) 	<p>M/s Larsen & Toubro Ltd. FG - Stores Kiads, Industrial Area Hebbal Hoota Galli, Mysore -570018 P.No.-0821-2405000 Fax No.-0821-2402462 Local Correspondence Address M/s Larsen & Toubro Ltd. Medical Equipments & Systems, D-2A, Prithviraj Road, C-Scheme, Jaipur P.No.-0141-2385921 Fax No, 0141-2373280</p>	Each	Model- Star 50, Brand - L&T	<p>Rs. 1,20,000.00 (One Lakh Twenty Thousand Only) + VAT@ 4% or as applicable</p> <p>CAMC for next 7 years after expiry of three years comprehensive guarantee period.</p> <p>CAMC charges 4200.00 IVth Year 4200.00 Vth Year 4200.00 VIth Year 4200.00 VIIth Year 4200.00 VIIIth Year 4200.00 IXth Year 4200.00 Xth Year</p>	<p>CAMC as above after three years guarantee period will be operative after a written request made by concerned purchase officers to the firm M/s Larsen & Toubro Ltd. FG - Stores Kiads, Industrial Area Hebbal Hoota Galli, Mysore</p>

GOVERNMENT OF RAJASTHAN
DIRECTORATE MEDICAL, HEALTH & W. SERVICES RAJASTHAN, JAIPUR

Item and rate approved for the rate contract period 2008-10 ending 30.9.2010 for Part-B (Equipments & Instruments) Section - M (FBNC & NBC) Item- Bilirubinometer

R.C. No. XXIV (Page No. 3)

R.C. No: DMHS/SPO/E&I/RC-XXIV/2008-10/09/ 1871


Dated: 4/6/09

S. No.	Name of the Item with Specifications	Name of Firm with Address	Packing Unit	Brand	Approved Rate Per Packing Unit	Remarks
1	2	3	4	5	6	7
	<p>Bilirubinometers are instruments that are simple to operate and are designed to measure the concentration of Bilirubin in blood.</p> <p>• <i>Operational requirements:</i> Unit should be Truly Portable Transcutaneous Billirubinometer for non- invasive measurement of bilirubin. Unit can be used during and after Phototherapy.</p> <p>MEASUREMENT PRINCIPLE: Reflectance photometry LIGHT SOURCE: Type - Xenon Flash Tube Bulb Life minimum 30000 measurement DETECTOR: Two Photocell Systems Accuracy: Imprecision (CV%) < 2% Correlation Between TcBI and Lab Values for serum bilirubin levels > 0.92 MEASURING RANGE 1 to 40 in Transcutaneous bilirubin index (TcBI) 10 to 400 in umol/l 0.58 to 23.5 mg/dL DISPLAY 2 digits LED MEASURING CYCLE TIME: Less than 4 seconds Three years comprehensive guarantee of whole unit</p>	<p>M/s H L Medical , System Pvt, Ltd. 101, Linkway Industrial Estate, Link Road, Malad (West) Mumbai- 400064 P.No.-022-66681995 Fax No.-02266681996</p>	Each	<p>Model - JM 103 Brand - Draeger Medical Germany</p>	<p>Rs. 1,50,000.00 (Rs. One Lakh Fifty Thousand only) + VAT/CST @4% or as applicable CAMC for next 7 years after expiry of three years comprehensive guarantee period. CAMC charges 3% Rs. 4500.00 IVth Year 3.3% Rs. 4950.00 Vth Year 3.6% Rs. 5400.00 VIth Year 3.9% Rs. 5850.00 VIIth Year 4.2% Rs. 6300.00 VIIIth Year 4.5% Rs. 6750.00 IXth Year 4.8% Rs. 7200.00 Xth Year</p>	<p>CAMC as above after three years guarantee period will be operative after a written request made by concerned purchase officers to the firm M/s H L Medical System Pvt, Ltd. 101, Linkway Industrial Estate, Link Road, Malad (West) Mumbai</p>

Director (PH)
Medical & Health Services
Rajasthan, Jaipur

<ul style="list-style-type: none">• ECG- 3 & 5 lead measurement and simultaneous monitoring of temperature• Monitor should be upgradeable to Dual IBP & EtCO2 at site• Accessories required:<ul style="list-style-type: none">• 3 and 5 lead ECG cable with disposable electrodes – 4& no of disposable electrodes• NIBP CUFF- Neonatal, Paediatric• Temperature probe• SPO2 PROBE – Three reusable sensors each for Neonate and Paediatric applications.• Monitor should have built in Electro Cautery & Defibrillator protection• Should work on 220 V AC \pm 10%, 50 Hz• US FDA approved <p>The complete system including all spare parts should be guaranteed for three years from the date of installation.</p>					
---	--	--	--	--	--

al


Director (PH)
Medical & Health Services
Rajasthan, Jaipur
11/11/09 31