

**GOVERNMENT OF RAJASTHAN
DIRECTORATE MEDICAL & HEALTH SERVICES RAJASTHAN**

No. SPO/E&I/App. List/2008-10/09/ 2777

Dated: 16/7/09

75

All Administrative Officers

**Sub: - Purchase of Store & Approval of Items Part-B (Equipments & Instruments)
Section-M (FBNC & NBC) Item- Mobile X-Ray Machine for the Rate Contract
& Supply period 2008-10 ending on 30-09-2010.**

In continuation to this Directorate letter no. SPO/E&I/App. List/2008-2010/08/2779 dated 22.10.08, No. 2799 dated 27-10-08, 3004 dated 25-11-08, 3209 dated 31.12.08, 382 dated 28-01-09, 478 dated 04-02-09, 558 dated 09-02-09, 567 dated 10-02-09, 623 Dated 13-02-09, 720 Dated 20-02-09, 987 dated 16.03.09, 1037 dated 18.3.09, 1038 dated 20.3.09, 1112 Dated 25-03-09, 1157 dated 31-03-09, 1353 dated 16-04-09, 1380 dated 20.4.09, 1381 dated 20.4.09, 1461 dated 27-04-09, 1571 dated 08-05-09, 1801 Dated 4-06-09, 1872 Dated 15-06-09, 1913 Dated 22-06-09 & 1974 Dated 02-07-09. Please find enclosed herewith the approved list showing the details such of the firm and rates etc. approved in respect of the firm and rates approved in respect of item Section-M (FBNC & NBC) Item- Mobile X-Ray Machine for the Rate contract the terms & conditions will remain same as already communicate to you earlier vide letter No. SPO/ E&I/App. List/ 2008-2010/ 08/ 2779 Dated 22-10-08 & Special Terms & Condition enclose herewith.

The supplier has executed the agreement bond. Therefore the payment of the supplier should be released to them, provided the goods are according to approved specification and rates, to your entire satisfaction as per rules.

It may please be ensured, if there is any recovery against the approved suppliers pertaining to the previous/ current rate contract on account of Liquidated Damages/ Risk Purchases/ Breakage/ Shortage/ Sub Standard etc. the same may please be recovered at your end before making payment.

Receipt of this communication may please be acknowledged.

Encl: - As above


Director (PH)

Medical & Health Services

Rajasthan, Jaipur

Dated: 16/7/09

No. SPO/E&I/App. List/2008-10/09/ 2007

Copy forwarded to the following for information & N/A:

1. PA To DM&HS (PH)/ (Aids)/(FW).
2. PA To Project Director, RHSDP/ NRHM.
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota & Udaipur.
4. Superintendent, SMS Hospital, Zanana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, TB & Chest Hospital, Jaipur.
5. Superintendent Associated Group Of Hospital-Ajmer, M.G. Hospital- Jodhpur, Ummaid Hospital- Jodhpur, PDM Hospital-Bikaner, MBS Hospital- Kota, & A.G. Hospital- Udaipur.
6. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj., Jaipur.
7. Addl. Director, Medical & Health Services (FW)/ (RH)/ (Hospital Admn.).
8. PA to FA of This Directorate along with 5 copies.
9. Joint Director, Medical & Health Services, (Zones) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharapour.
10. All In-charge TB Clinic's
11. All Dy. CM&HO's
12. In-charge Server Room, Head Quarter
13. Concerned Firms.
14. Office Order File.


Chief Accounts Officer & Secretary,

Store Purchase Organisation,

Medical & Health Directorate,

Rajasthan Jaipur

GOVERNMENT OF RAJASTHAN
DIRECTORATE MEDICAL, HEALTH & W. SERVICES RAJASTHAN, JAIPUR

Item and rate approved for the rate contract period 2008-10 ending 30.9.2010 for Part-B (Equipments & Instruments) Section -M (FENC & NBC) Item- Mobile X-Ray Machine

R.C. No: DMHS/SPO/E&I/RC-XXXI/2008-10/09/ 2007


R.C. No. XXXI (Page No. 2)
 Dated: 16/12/2007


S. No.	Name of the Item with Specifications	Name of Firm with Address	Packing Unit	Brand	Approved Rate Per Packing Unit	Remarks
1	2	3	4	5	6	7
1.	<p>Mobile X-Ray Machine:-</p> <ul style="list-style-type: none"> • a. The Unit should be able to give an high high frequency up to 70ma. • b. The Unit should have 40 to 100 kV. • c. X ray Tube DSA type - Diagnostic Stationary Anode type. • d. The Electronic Timer is to have latest Microprocessor based technology or should have the MAS Integration. • System should have online monitoring of KV and MAS ensuring optimum usage of the X-ray tube. • The Unit should work on high frequency X-ray generation technology & it should be able to work with input voltage of 170-250 volts, with mains resistance ≤ 1.5 Ohm. Overload protection by continuous monitoring of parameters. • The Collimator in the Equipment should have high intensity light for clear visualization of the area of interest with auto shut facility to ensure longer life of the collimator lamp. • The equipment should have Cassette storage Box for storage of cassettes and films. • i. The Equipment should not weigh more than 250 Kgs. • The Height to the Unit (the column stand) should be less than 160 cms. • Feed back mechanism for error identification available on the equipment. 	<p>M/s Allengers Medical System Ltd. S.C.O. 212-214, Sector 34-A Chandigarh-160022 Phone No. 0172-3012280-E4 Fax No.- 0172-2621912-13</p>	Each	<p>MARS 3.5 SBM. (70ma, 110kv high frequency X-Ray machine with spring balanced mobile stand), Brand - India</p>	<p>Rs. 1,59,000/- (15.90 Lakh) + 15% GST or as applicable CAMC for next 7 years after expiry of three years comprehensive guarantee period. CAMC charges 4,960.00 IVth Year 4,960.00 Vth Year 4,960.00 VIth Year 4,960.00 VIIth Year 4,960.00 VIIIth Year 4,960.00 IXth Year 4,960.00 Xth Year</p>	<p>(1) CAMC as above after three years guarantee period will be operative after a written request made by concerned purchase officers to the firm M/s Allengers Medical System Ltd. S.C.O. 212-214; Sector 34-A Chandigarh-160022 (2) Note- Purchase officer are required to ensure both auto and manual mode operation are available in the mobile X-ray machine at the time of delivery.</p>

101

102

Contd.

S. No.	Name of the item with Specifications	Name of Firm with Address	Packing Unit	Brand	Approved Rate Per Packing Unit	Remarks
1	2	3	4	5	6	7
	<ul style="list-style-type: none"> • Manufacturer must have been successfully manufacturing and marketing the quoted model for the last three years. • The equipment should be certified by BIS/CE & AERB for Safety regulation & should have ISO certificate. Enclose required valid certificate along with tender documents. • Unit should have electronic feather touch key board. • Both auto and manual mode selection should be possible. • Should have independent Kvp & mAs Display. • Audible X ray ON alarm should be available for safety. • Spring balanced tube stand with rotation in all direction having horizontal and vertical travel arms for better positioning. • The system should switch to standby mode automatically if unit is idle for more than 5-7 minutes. • Unit should provide with main power cable having length not less than 3 meters. • Unit must have electronic timer 0.02 to 3 sec. • Out put of the X ray generator should be 2.5 KW or more. • Movements are: <ol style="list-style-type: none"> i. Up/Down ii. Vertical travel iii. In/out travel iv. Angular travel • Three years comprehensive guarantee of whole machine • CAMC for next seven years with spares and parts 					


 Director (PH)
 Medical & Health Services
 Rajasthan, Jaipur


 27/10/11