

GOVERNMENT OF RAJASTHAN
DIRECTORATE MEDICAL & HEALTH SERVICES RAJASTHAN

No. SPO/E&I/App. List/2008-10/09/ 567

Dated: 10/2/09

All Administrative Officers

Sub :- Purchase of Store & Approval of Items (Equipments & Instruments) Part-B, Section-B for the Rate contract period 2008-10 ending on 30-09-2010 (B/43- Heat Warner & 44 - Phototherapy source unit with infant care trolley)

In continuation to this Directorate letter no. SPO/E&I/App. List/2008-2010/08/2779 dated 22.10.08, No. 2799 dated: 27-10-08, 3004 dated 25-11-08, 3209 dated 31.12.08, 382 Dated 28-01-09 & 478 Dated 04-02-09. Please find enclosed herewith the approved list showing the details such of the firm and rates etc. approved in respect of the firm and rates approved in respect of item (B/43- Heat Warner & 44 - Phototherapy source unit with infant care trolley). The terms & conditions will remain same as already communicate to you earlier vide letter No. SPO/E&I/App. List/2008-2010/08/2779 Dated: 22-10-08.

The supplier has executed the agreement bond. Therefore the payment of the supplier should be released to them, provided the goods are according to approved specification and rates, to your entire satisfaction as per rules.

It may please be ensured, if there is any recovery against the approved suppliers pertaining to the previous/ current rate contract on account of Liquidated Damages/ Risk Purchases/ Breakage/ Shortage/ Sub Standard etc. the same may please be recovered at your end before making payment.

Receipt of this communication may please be acknowledged.

Encl: - As above

Director (PH)
Medical & Health Services
Rajasthan, Jaipur

Dated: 10/2/09

No. SPO/E&I/App. List/2008-10/09/ 567

Copy forwarded to the following for information & N/A:

1. PA To DM&HS (PH)/ (Aids)/(FW).
2. PA To Project Director, RHSDP/ NRHM.
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota & Udaipur.
4. Superintendent, SMS Hospital, Zanana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, TB & Chest Hospital, Jaipur.
5. Superintendent Associated Group Of Hospital-Ajmer, M.G. Hospital- Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital- Kota, & A.G. Hospital-Udaipur.
6. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj., Jaipur.
7. Addl. Director, Medical & Health Services (FW)/ (RH)/ (Hospital Admn.).
8. PA to FA of This Directorate.
9. Joint Director, Medical & Health Services, (Zones) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
10. All In-charge T.B Clinic's
11. All Dy. CM&HO's
12. In-charge Server Room, Head Quarter
13. Concerned Firms.
14. Office Order File.

Chief Accounts Officer & Secretary,
Store Purchase Organisation,
Medical & Health Directorate,
Rajasthan Jaipur

GOVERNMENT OF RAJASTHAN
DIRECTORATE MEDICAL, HEALTH F.W. SERVICES RAJASTHAN, JAIPUR

Item and rate approved for the rate contract period 2008-10 ending 30.9.2010 for Part-B (Equipments & Instruments) Section-B (Gynecological Equipment & Instrument)

R.C. No. IX, (Page No. 2)


R.C. No: DMHS/SPO/E&I/RC-IX/2008-10/09/ 567

Dated: 10/2/09

S. No.	Cat. No.	Name of the Item with Specifications	Name of Firm with Address	Packing Unit	Brand	Approved Rate
1	2	3	4	5	6	7
1	B/43	Heat Warmer Servo Controlled: Control Modes: i. Skin control, ii. Air Control, iii. Manual control range of temp. For skin control: 34 to 38 degree C. range of temp. for air control : room temp. to 39 degree C. Manual mode: i. Time programmable, ii. Heater out put settable, iii. Over temp. Cut off Heater: a. Ceramic Type. b. Box can be swiveled to take X-Ray. Baby Tray (Stainless Steel) a. Clear acrylic collapsible side panels. b. Stainless Steel baby tray with transparent slot to incorporate the facility of under surface phototherapy. c. Acrylic and cushion mattress with headup/ down facility. Acrylic X-ray guide for positioning the cassette. e. Strong hinges for side panels. Lower shelf and 3 drawer storages cabinet Side mounted monitor tray pole for syringe pump/ infusion pump IV stand It should have long display for Air & Baby. There should be selectable arrangement to select temp. Range in. Should have LCD Display for programming different feature. Feather touch key pad. RS 232 connecting point. Voice calling system after selective time is over. Display: a. Skin temperature, b. Air temperature, c. Heater power (% Heater output on LCD display) Alarms: I. High + 1.0 degree. C, ii. Low - 1.0 degree. C b. Skin mode I. High + 1.0 degree. C, i. Low - 1.0 degree. C Power failure. Safety cut off at 39 degree. C power source 230 volt. 50 Hz.	M/s Meditrix Instruments, Madhuban, 98 Pathare Wadi, Opp. Bageecha Restaurant, Malvani Church, Marve Road, Malad(W), Mumbai- 400095. P.No.-022-28616463, 28633878 Fax No. 022-28620176	Each	Meditrix Instruments Own Make	Rs. 30,000/- (Rs. Thirty thousand only) + 12.5% Tax extra or as applicable.

R02
9/2
9/2

2	B/44	<p>Phototherapy source Single surface unit with infant care trolley. Unit should have Six 2 Feet Tubular. 4 Light blue radiations (TL 52) & 2 Lights should be white of reputed co. Unit should have in built silent cooling Fan. Height adjustable facility. Easy swiveling for angel adjustment, when stand is to be used in conjunction Radiant Warmer. Clear Filter. Display total hours of tube light life up to 2000 hours and more. Unit should have a non conductive tube box to avoid any kind of electric shock hazard to patient or operator. In built digital time hour meter with reset facility for both under & overheat unit & should also measure the total exposure timings. Frequency 420 to 500nm. Should have baby tray made up of good quality sturdy transparent polycarbonate/ acrylic sheet with folding side support Accessible from all sides. Should have all utility timers.</p>	<p>M/s Meditrin Instruments, Madhuban, 98 Pathare Wadi, Opp. Bageecha Restaurant, Malvani Church, Marve Road, Malad(W), Mumbai- 400095. P.No.-022-28616463, 28633878 Fax No. 022-28620176</p>	<p>Each</p>	<p>Meditrin Instruments Own Make</p>	<p>Rs. 17,100/- (Rs. Seventeen thousand One hundred only) + 12.5% Tax extra or as applicable.</p>
---	------	---	---	-------------	--	---



RO
9/2

Director (PH)
Medical & Health Services
Rajasthan, Jaipur
9/2